

Customer:

Recreational8 LLC
7173 Lake Worth Rd
Lake Worth, FL 33467

Sample ID:

675PR

Laboratory Number: ATL-4061



Report Issue Date
1/19/2022

Extraction Technician: HP
Analytical Chemist: HP

Sample Description/Size:

D8 Flower Lifter

Unit Weight: 2.5g



Hernan Prieto

CANNABINOID PROFILE

Order Date
1/13/2022

Analysis Date
1/18/2022

Cannabinoids (HPLC)

Results

Cannabinoid (%)

Test	LOD (mg/g)	mg/g	%	0	20	40	60	80	100
Cannabidiol (CBD)	<0.05	15.93	1.59						
Cannabidiol Acid (CBD-A)	<0.04	131.59	13.16						
Cannabigerol (CBG)	<0.05	0	0						
Cannabigerol Acid (CBG-A)	<0.04	2.65	0.27						
Cannabidivarin (CBDV)	<0.05	0	0						
Delta 9-Tetrahydrocannabinol (THC)	<0.05	2.17	0.22						
Delta 8-Tetrahydrocannabinol	<0.05	35.33	3.53						
Cannabichromene (CBC)	<0.05	1.33	0.13						
Cannabichromene Acid (CBC-A)	<0.05	6.49	0.65						
Delta-9-Tetrahydrocannabinolic Acid (THC-A)	<0.04	3.77	0.38						

Cannabinoids Total

Test	mg/g	%	0	20	40	60	80	100
Max Active THC	5.94	0.59						
Max Active CBD	147.52	14.75						
T.Active Cannabinoids	199.26	19.93						
Total Cannabinoids	219.52	21.95						

Analysis Method: ATL-PLC-001h

Following USDA guidelines on uncertainty, Accurate Test Lab is uncertainty are calculated for CBDa and CBD at +/- 4%. The uncertainty for THCa and THC are +/- 5%. This implies the range for a 10% value of CBD to be 9.6-10.4%. The uncertainty range for a 0.30% value of THC would be 0.28-0.32%. The measurement uncertainty is calculated using a coverage factor of 2.

Reporting Limits will vary based on sample extraction weight used for the analysis. Accurate Test Lab, LLC utilizes based upon traceable Reference Standards and Certified Reference Material to calibrate analytical instruments along with proven analytical methods. The methods are applied in the most ethical manner following good laboratory practice guidelines. The results of this report are based solely on the sample submitted and cannot be reproduced. Results only apply to samples within COA as received. Certificate of Analysis shall not be reproduce except in full without approval of Accurate Test Lab, LLC.

N/D = Not Detected

